

## Synthetic Natural Stretch Vinyl Powder Free Exam Gloves



SYNTHETIC

**Product Type:** Powder free ambidextrous non-sterile examination glove

Product Re-order	Small	Medium	Large	XLarge
Product code	GL6621	GL6622	GL6623	GL6625

**Material specification:** Polyvinyl Chloride (PVC)  
**Manufacturing process types:** Straight vinyl dip with polyurethane coating  
**Colour:** Natural  
**Texture and grip:** Non textured surface  
**Former type:** Non textured former

**Freedom from holes BS EN 455-1**

<b>Water Tight Test</b>	General Inspection Level 1, AQL 1.5
-------------------------	-------------------------------------

**Physical properties BS EN 455-2** N=13, Median Value, Dumbbell sample width 3mm

Dimensions	Small	Medium	Large	XLarge
<b>Palm width (mm)</b>	80 ± 10	95 ± 10	110 ± 10	≥110
<b>Length (mm)</b>	240 minimum	240 minimum	240 minimum	240 minimum
<b>Thickness (mm) palm</b>	0.09 ± 0.02	0.09 ± 0.02	0.09 ± 0.02	0.09 ± 0.02
<b>Thickness (mm) finger</b>	0.07 ± 0.02	0.07 ± 0.02	0.07 ± 0.02	0.07 ± 0.02

Force At Break Test	Throughout Shelf Life	Small	Medium	Large	XLarge
		Min. 3.6 Newton	Min. 3.6 Newton	Min. 3.6 Newton	Min. 3.6 Newton

**Residual Powder ASTM D6124-01**

<b>Powder Content</b>	Samples should contain no more than 2.0mg of any filterable residue per glove
-----------------------	---

<b>Food Safe</b>	Food safe gloves (EC 1935/2004 and European Regulation 10/2011, except fatty foods).
------------------	--

**Quality Standards (ISO BS EN) Expectation**

**Manufacturing QMS:** BS EN ISO 9001 / BS EN ISO 13485  
**Sampling:** ISO 2859-1  
**Medical Device Directive:** 93/42/EEC and subsequent amendments. EN455 series of standards. CE Class I Medical Device.  
**Personal Protective Equipment Directive:** 89/686/EEC. EN420 and EN374 series of standards. CE 0086 Category III  
**Technical File Requirements & Availability:** Available for inspection by Brosch direct or their notified body  
**Shelf life:** 3 years

Size	Inner Barcode	Outer Barcode
<b>Small</b>	5024951003182	5024951003199
<b>Medium</b>	5024951003205	5024951003212
<b>Large</b>	5024951003229	5024951003236
<b>XLarge</b>	5024951003243	5024951003250

Packaging Size	
<b>Inner Dispenser</b>	230mm (L) x 120mm (W) x 63mm (H)
<b>Outer Carton</b>	330mm (L) x 248mm (W) x 240mm (H)

  

Quantities	
<b>Inner Dispenser</b>	100 Gloves per Dispenser. Sample N=8. Acceptance criteria: ≥ 98% average quantity
<b>Outer Carton</b>	10 Dispensers per Carton
<b>Pallet</b>	84 cases

**Lot Numbering Requirements:** Brosch direct standard LOT numbering system

**Testing Requirements & Frequency:**

Each size and LOT will be tested by the manufacturer to BS EN 455. Shipment release will be authorised by Brosch direct QA department. Independent tests will be carried out in accordance with the Brosch direct statistical sampling plan.

Brosch direct Head Office Technical Contact Details	
<b>Contact</b>	Karen Gunning
<b>Address</b>	South Fen Road, Bourne, Lincolnshire, PE10 0DN
<b>Tel no.</b>	01733 405 581
<b>Email</b>	k.gunning@broschdirect.com

### Glove Symbols



Latex Free



Powder Free



Class 1 Medical Device  
Manufactured to BS EN 455



0086  
CAT III PPE



EN374-3: Chemical Hazards



EN374-2: Micro Organism Hazards



Single Use



Lot/Batch Number



Use By Date



Date of Manufacture



Manufactured By



Packaging Recyclable



Storage Temperature



Avoid Direct Sunlight/  
Fluorescent Light



Avoid Damp Storage